

Fluoroscopy skin injuries

Avoid exposing the same area of the skin by means of different projections.

Vary the beam entrance port by rotating the tube around the patient.

Oblique projections increase the Entrance Skin Dose (ESD).

Avoid the use of magnification.

Example of skin injury as result of long screening procedure (Wagner, 2007).

